

**P.2.14.2 CSF 14-3-3 Automated Capillary Western Assay**

Studies	Design	Total N	Sens (95%CI)	Spec (95%CI)	Measure	Summary of findings (95%CI)	Risk of bias	Inconsistency	Indirectness	Imprecision	Other considerations	Quality
<b>SECONDARY CARE</b>												
1 study (Fourier 2017)	Retrospective	268	0.94 (0.85, 0.97)	0.95 (0.91, 0.98)	LR+	19.84 (10.46, 37.65)	Not serious	n/a	Not serious	Not serious	-	HIGH
					LR-	0.07 (0.03, 0.16)	Not serious	n/a	Not serious	Not serious		HIGH